

## 9843 FAQ'S

### **Q. What is included in the 9843 package assembly?**

- A.** • 9843 Pulse Oximeter • 9843 Manual on CD • 9840SA Co2 Sensor  
• 8000AA Pulse Oximeter Sensor • 6 AA-Size Alkaline Batteries

### **Q. How do you know you're getting an accurate reading?**

**A.** The tricolor LED display (m) provides a visual indication of the pulse signal quality, while blinking at the corresponding pulse rate. This display changes colors to alert you to changes in pulse quality that may affect the readings: *green* indicates a good pulse signal, *yellow* indicates a marginal pulse signal, and *red* indicates an inadequate pulse signal. Allow the 9843 to stabilize, observing about 10 seconds of continuous green-colored pulse quality before relying on the displayed values. It is common for the displayed values to fluctuate slightly over a period of several seconds. If the pulse quality display blinks yellow or red, try another finger

### **Q. What kind of battery is used in the 9843?**

**A.** Six AA Alkaline batteries.

### **Q. How many hours of battery life can I expect?**

**A.** The 9843 will typically operate for 90 hours continuously between alkaline battery replacements for pulse oximetry alone, or 20 hours continuously with both Co2 detection and pulse oximetry operating concurrently.

### **Q. How do you access the memory stored in device?**

**A.** NONIN's nVISION data management software and 1000MC serial cable will allow you to connect the 9843 to PC for downloading data and view data.

### **Q. What happens to the stored memory after 72 hours have been displayed?**

**A.** The memory in the device functions as an "endless loop." When the memory fills up, the unit begins overwriting the oldest data with the newest.